

Nominal <i>(Nominal)</i>	Patrón <i>(Standard)</i>	UBP <i>(UUT)</i>	Error <i>(Error)</i>	Exactitud <i>(Accuracy)</i>	Incertidumbre <i>(Uncertainty)</i>
30 °C	30.263 °C	30.5 °C	0.237 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	30.570 %HR	35.6 %HR	5.030 %HR	± 5.0 %HR	± 1.9E+00 %HR
50 %HR	50.355 %HR	54.6 %HR	4.245 %HR	± 5.0 %HR	± 1.9E+00 %HR
70 %HR	70.448 %HR	74.4 %HR	3.952 %HR	± 5.0 %HR	± 1.9E+00 %HR