

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
30 °C	30.263 °C	30.6 °C	0.337 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	30.570 %HR	35.7 %HR	5.130 %HR	± 5.0 %HR	± 1.9E+00 %HR
50 %HR	50.355 %HR	55.2 %HR	4.845 %HR	± 5.0 %HR	± 1.9E+00 %HR
70 %HR	70.448 %HR	73.7 %HR	3.252 %HR	± 5.0 %HR	± 1.9E+00 %HR