

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
25 °C	24.848 °C	25.0 °C	0.152 °C	± 0.5 °C	± 4.0E-01 °C
# 25 °C	24.848 °C	24.7 °C	-0.148 °C	± 0.5 °C	± 7.4E-02 °C
# 50 °C	49.917 °C	49.9 °C	-0.017 °C	± 0.5 °C	± 7.4E-02 °C
# 70 °C	69.972 °C	70.1 °C	0.128 °C	± 0.5 °C	± 7.4E-02 °C
25 %HR	23.413 %HR	24.4 %HR	0.987 %HR	± 3 %HR	± 1.9E+00 %HR
50 %HR	48.975 %HR	49.7 %HR	0.725 %HR	± 3 %HR	± 1.9E+00 %HR
70 %HR	69.425 %HR	70.1 %HR	0.675 %HR	± 3 %HR	± 1.9E+00 %HR