

Clase de Exactitud (Accuracy Class): Clase I (Class I)

Valor UBP	Valor Patrón	Error	EMP	Incertidumbre (Uncertainty)		Distancia de objetos (Colocados)	Distancia detectada		Error	Incertidumbre (Uncertainty)
m	m	mm	mm			m	m	mm	mm	
0 a 1	0.99965	0.350	± 0.4	±	4.3E-01 mm	1	1.00029	0.00029	4.3E-01	mm
0 a 2	2.00015	-0.150	± 0.6	±	4.8E-01 mm	2	2.00070	0.00070	4.8E-01	mm
0 a 3	2.99962	0.380	± 0.8	±	4.8E-01 mm	3	3.00021	0.00021	4.8E-01	mm
0 a 4	3.99988	0.120	± 1	±	4.8E-01 mm	4	4.00045	0.00045	4.8E-01	mm
0 a 5	4.99979	0.210	± 1.2	±	4.8E-01 mm	5	5.00025	0.00025	4.8E-01	mm
0 a 6	6.00026	-0.260	± 1.4	±	3.6E-01 mm	6	6.00055	0.00054	3.6E-01	mm
0 a 7	7.00007	-0.070	± 1.6	±	3.6E-01 mm	7	7.00043	0.00043	3.6E-01	mm
0 a 8	7.99980	0.200	± 1.8	±	3.6E-01 mm	8	8.00022	0.00022	3.6E-01	mm
0 a 9	9.00029	-0.290	± 2	±	3.6E-01 mm	9	9.00018	0.00018	3.6E-01	mm
0 a 10	9.99915	0.850	± 2.2	±	3.6E-01 mm	10	9.98500	-0.01500	3.6E-01	mm
0 a 11	11.00015	-0.150	± 2.4	±	3.6E-01 mm	11	11.00027	0.00027	3.6E-01	mm
0 a 12	12.00048	-0.480	± 2.6	±	3.6E-01 mm	12	12.00085	0.00085	3.6E-01	mm
0 a 13	13.00037	-0.370	± 2.8	±	3.6E-01 mm	13	13.00059	0.00059	3.6E-01	mm
0 a 14	14.00043	-0.430	± 3	±	3.6E-01 mm	14	14.00057	0.00057	3.6E-01	mm
0 a 15	15.00052	-0.520	± 3.2	±	3.6E-01 mm	15	15.00066	0.00066	3.6E-01	mm