

Clase de Exactitud (Accuracy Class): Clase I (Class I)

Valor UBP	Valor Patrón	Error	EMP	Incertidumbre (Uncertainty)		Valor UBP	Valor Patrón	Error	EMP	Incertidumbre (Uncertainty)	
m	m	mm	mm			m	m	mm	mm		
0 a 1	0.99965	0.350	± 0.4	±	4.3E-01 mm						
0 a 2	2.00015	-0.150	± 0.6	±	4.8E-01 mm						
0 a 3	2.99962	0.380	± 0.8	±	4.8E-01 mm						
0 a 4	3.99988	0.120	± 1	±	4.8E-01 mm						
0 a 5	4.99979	0.210	± 1.2	±	4.8E-01 mm						
0 a 6	6.00026	-0.260	± 1.4	±	3.6E-01 mm						
0 a 7	7.00007	-0.070	± 1.6	±	3.6E-01 mm						
0 a 8	7.99980	0.200	± 1.8	±	3.6E-01 mm						
0 a 9	9.00029	-0.290	± 2	±	3.6E-01 mm						
0 a 10	9.99915	0.850	± 2.2	±	3.6E-01 mm						
0 a 11	11.00015	-0.150	± 2.4	±	3.6E-01 mm						
0 a 12	12.00048	-0.480	± 2.6	±	3.6E-01 mm						
0 a 13	13.00037	-0.370	± 2.8	±	3.6E-01 mm						
0 a 14	14.00043	-0.430	± 3	±	3.6E-01 mm						
0 a 15	15.00052	-0.520	± 3.2	±	3.6E-01 mm						