

Nominal <i>(Nominal)</i>	Patrón <i>(Standard)</i>	UBP <i>(UUT)</i>	Error <i>(Error)</i>	Exactitud <i>(Accuracy)</i>	Incertidumbre <i>(Uncertainty)</i>
10 °C	10.519 °C	10.6 °C	0.081 °C	± 0.5 °C	± 4.0E-01 °C
32 °C	32.168 °C	31.2 °C	-0.968 °C	± 0.5 °C	± 4.0E-01 °C
40 %HR	39.670 %HR	45.7 %HR	6.030 %HR	± 3.0 %HR	± 1.9E+00 %HR
75 %HR	75.100 %HR	80.3 %HR	5.200 %HR	± 3.0 %HR	± 1.9E+00 %HR