

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
10 °C	10.519 °C	10.5 °C	-0.019 °C	± 0.5 °C	± 4.0E-01 °C
32 °C	32.168 °C	31.1 °C	-1.068 °C	± 0.5 °C	± 4.0E-01 °C
40 %HR	39.670 %HR	45.4 %HR	5.730 %HR	± 3.0 %HR	± 1.9E+00 %HR
75 %HR	75.100 %HR	79.8 %HR	4.700 %HR	± 3.0 %HR	± 1.9E+00 %HR