

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
10 °C	10.138 °C	9.9 °C	-0.238 °C	± 0.5 °C	± 4.0E-01 °C
32 °C	32.569 °C	32.9 °C	0.331 °C	± 0.5 °C	± 4.0E-01 °C
40 %HR	40.693 %HR	42.7 %HR	2.007 %HR	± 3.0 %HR	± 1.9E+00 %HR
75 %HR	75.100 %HR	77.8 %HR	2.700 %HR	± 3.0 %HR	± 1.9E+00 %HR