

Nominal <i>(Nominal)</i>	Patrón <i>(Standard)</i>	UBP <i>(UUT)</i>	Error <i>(Error)</i>	Exactitud <i>(Accuracy)</i>	Incertidumbre <i>(Uncertainty)</i>
10 °C	10.519 °C	10.8 °C	0.281 °C	± 0.5 °C	± 4.0E-01 °C
32 °C	32.028 °C	32.4 °C	0.372 °C	± 0.5 °C	± 4.0E-01 °C
40 %HR	39.670 %HR	40.8 %HR	1.130 %HR	± 3.0 %HR	± 1.9E+00 %HR
75 %HR	75.100 %HR	75.9 %HR	0.800 %HR	± 3.0 %HR	± 1.9E+00 %HR