

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
10 °C	10.519 °C	10.8 °C	0.281 °C	± 0.5 °C	± 4.0E-01 °C
32 °C	32.028 °C	32.2 °C	0.172 °C	± 0.5 °C	± 4.0E-01 °C
40 %HR	39.670 %HR	41.4 %HR	1.730 %HR	± 3.0 %HR	± 1.9E+00 %HR
75 %HR	75.100 %HR	76.8 %HR	1.700 %HR	± 3.0 %HR	± 1.9E+00 %HR