

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
0 °C	0.381 °C	0.3 °C	-0.081 °C	± 0.5 °C	± 4.0E-01 °C
5 °C	5.545 °C	5.3 °C	-0.245 °C	± 0.5 °C	± 4.0E-01 °C
10 °C	10.138 °C	10.1 °C	-0.038 °C	± 0.5 °C	± 4.0E-01 °C