

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
0 °C	0.381 °C	0.5 °C	0.119 °C	± 0.5 °C	± 4.0E-01 °C
5 °C	5.545 °C	5.5 °C	-0.045 °C	± 0.5 °C	± 4.0E-01 °C
10 °C	10.138 °C	10.0 °C	-0.138 °C	± 0.5 °C	± 4.0E-01 °C