
Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
0 °C	0.481 °C	0.7 °C	0.219 °C	± 0.5 °C	± 4.0E-01 °C
5 °C	4.953 °C	5.1 °C	0.147 °C	± 0.5 °C	± 4.0E-01 °C
10 °C	10.509 °C	10.2 °C	-0.309 °C	± 0.5 °C	± 4.0E-01 °C