

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
-15 °C	-15.493 °C	-15.6 °C	-0.107 °C	± 0.5 °C	± 4.0E-01 °C
-25 °C	-25.972 °C	-25.6 °C	0.372 °C	± 0.5 °C	± 4.0E-01 °C