

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
0 °C	0.351 °C	0.4 °C	0.049 °C	± 0.5 °C	± 4.0E-01 °C
5 °C	5.545 °C	5.8 °C	0.255 °C	± 0.5 °C	± 4.0E-01 °C
10 °C	10.138 °C	9.9 °C	-0.238 °C	± 0.5 °C	± 4.0E-01 °C
57 %HR	57.595 %HR	60.8 %HR	3.205 %HR	± 3.0 %HR	± 1.9E+00 %HR