

<b>Nominal</b> ( <i>Nominal</i> )	<b>Patrón</b> ( <i>Standard</i> )	<b>UBP</b> ( <i>UUT</i> )	<b>Error</b> ( <i>Error</i> )	<b>Exactitud</b> ( <i>Accuracy</i> )	<b>Incertidumbre</b> ( <i>Uncertainty</i> )
30 °C	30.032 °C	30.5 °C	0.468 °C	± 0.5 °C	± 4.0E-01 °C
35 %HR	34.149 %HR	32.0 %HR	-2.149 %HR	± 3.0 %HR	± 1.9E+00 %HR
50 %HR	50.100 %HR	48.1 %HR	-2.000 %HR	± 3.0 %HR	± 1.9E+00 %HR
75 %HR	75.427 %HR	72.9 %HR	-2.527 %HR	± 3.0 %HR	± 1.9E+00 %HR