

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
30 °C	30.313 °C	25.0 °C	-5.313 °C	± 1.0 °C	± 4.0E-01 °C
30 %HR	30.181 %HR	18.0 %HR	-12.181 %HR	± 3.0 %HR	± 1.9E+00 %HR
50 %HR	50.509 %HR	44.0 %HR	-6.509 %HR	± 3.0 %HR	± 1.9E+00 %HR
70 %HR	70.223 %HR	60.0 %HR	-10.223 %HR	± 3.0 %HR	± 1.9E+00 %HR