Patrón (Standard)		UBP (UUT)		Error (Error)	E.M.P (Tolerance)	Incertidumbre (Uncertainty)	
0.0625	in	0.062	in	-0.0005 in	$\pm 0.0001$ in	±	5.8E-04 in
0.1	in	0.100	in	0 in	$\pm 0.0001 in$	±	5.8E-04 in
0.125	in	0.125	in	0 in	$\pm 0.0001 in$	±	5.8E-04 in
0.2	in	0.200	in	0 in	$\pm 0.0001 in$	±	5.8E-04 in
0.25	in	0.250	in	0 in	$\pm 0.0001 in$	±	5.8E-04 in
0.3	in	0.300	in	0 in	$\pm 0.0001 in$	±	5.8E-04 in
0.35	in	0.350	in	0 in	$\pm 0.0001 in$	±	5.8E-04 in
0.4	in	0.400	in	0 in	$\pm 0.0001 in$	±	5.8E-04 in
0.425	in	0.425	in	0 in	$\pm 0.0001 in$	±	5.8E-04 in
0.45	in	0.450	in	0 in	$\pm 0.0001$ in	±	5.8E-04 in
0.5	in	0.500	in	0 in	$\pm 0.0001 in$	±	5.8E-04 in