

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
30 °C	29.908 °C	30.1 °C	0.192 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	28.760 %HR	29.7 %HR	0.940 %HR	± 3.0 %HR	± 1.9E+00 %HR
50 %HR	51.061 %HR	52.1 %HR	1.039 %HR	± 3.0 %HR	± 1.9E+00 %HR
70 %HR	72.196 %HR	73.3 %HR	1.104 %HR	± 3.0 %HR	± 1.9E+00 %HR