

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
30 °C	29.908 °C	29.9 °C	-0.008 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	28.760 %HR	30.6 %HR	1.840 %HR	± 3.0 %HR	± 1.9E+00 %HR
50 %HR	51.061 %HR	53.5 %HR	2.439 %HR	± 3.0 %HR	± 1.9E+00 %HR
70 %HR	72.196 %HR	72.7 %HR	0.504 %HR	± 3.0 %HR	± 1.9E+00 %HR