

Patrón <i>(Standard)</i>	UBP <i>(UUT)</i>	Error <i>(Error)</i>	Exactitud <i>(Accuracy)</i>	Incertidumbre <i>(Uncertainty)</i>	
0.29 RPM	0.3 RPM	0.01 RPM	± 0.29 RPM	±	1.0E-01 RPM
0.99 RPM	1 RPM	0.01 RPM	± 0.99 RPM	±	1.3E-01 RPM
4.99 RPM	5 RPM	0.01 RPM	± 4.99 RPM	±	1.0E-01 RPM
9.9 RPM	10 RPM	0.1 RPM	± 9.9 RPM	±	1.0E-01 RPM
19.9 RPM	20 RPM	0.1 RPM	± 20 RPM	±	1.0E-01 RPM
29.9 RPM	30 RPM	0.1 RPM	± 29.9 RPM	±	1.0E-01 RPM
49.8 RPM	50 RPM	0.2 RPM	± 49.8 RPM	±	1.0E-01 RPM
59.8 RPM	60 RPM	0.2 RPM	± 59.8 RPM	±	1.0E-01 RPM
99.8 RPM	100 RPM	0.2 RPM	± 99.8 RPM	±	8.2E-02 RPM
199.7 RPM	200 RPM	0.3 RPM	± 199.7 RPM	±	8.2E-02 RPM