

Patrón <i>(Standard)</i>	UBP <i>(UUT)</i>	Error <i>(Error)</i>	Exactitud <i>(Accuracy)</i>	Incertidumbre <i>(Uncertainty)</i>	
19.9 RPM	20 RPM	0.1 RPM	± 2 RPM	±	1.0E-01 RPM
29.9 RPM	30 RPM	0.1 RPM	± 2 RPM	±	1.0E-01 RPM
49.8 RPM	50 RPM	0.2 RPM	± 2 RPM	±	1.0E-01 RPM
59.8 RPM	60 RPM	0.2 RPM	± 2 RPM	±	1.0E-01 RPM
99.8 RPM	100 RPM	0.2 RPM	± 2 RPM	±	8.2E-02 RPM
199.7 RPM	200 RPM	0.3 RPM	± 2 RPM	±	8.2E-02 RPM

