

---

<b>Patrón (Standard)</b>	<b>UBP (UUT)</b>	<b>Error (Error)</b>	<b>Exactitud (Accuracy)</b>	<b>Incertidumbre (Uncertainty)</b>	
19.9 RPM	20 RPM	0.1 RPM	± 2 RPM	±	1.0E-01 RPM
29.9 RPM	30 RPM	0.1 RPM	± 2 RPM	±	1.0E-01 RPM
49.8 RPM	50 RPM	0.2 RPM	± 2 RPM	±	1.0E-01 RPM
59.8 RPM	60 RPM	0.2 RPM	± 2 RPM	±	1.0E-01 RPM
99.8 RPM	100 RPM	0.2 RPM	± 2 RPM	±	8.2E-02 RPM
199.7 RPM	200 RPM	0.3 RPM	± 2 RPM	±	8.2E-02 RPM