

Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	E.M.P (<i>Tolerance</i>)	Incertidumbre (<i>Uncertainty</i>)	
# 0.0 μm	0.0 μm	0.0 μm	± 2.5 μm	±	5.8E-02 μm
24.1 μm	24.2 μm	0.1 μm	± 2.5 μm	±	5.8E-02 μm
53.5 μm	54 μm	0.2 μm	± 2.5 μm	±	5.8E-02 μm
126.0 μm	126 μm	0.0 μm	± 3.8 μm	±	6.1E-02 μm
255.0 μm	255 μm	0.0 μm	± 7.7 μm	±	7.0E-02 μm
493.0 μm	493 μm	0.0 μm	± 15 μm	±	9.5E-02 μm