

| Patrón (Standard) | UBP (UUT) | Error (Error) | E.M.P (Tolerance) | Incertidumbre (Uncertainty) |            |
|-------------------|-----------|---------------|-------------------|-----------------------------|------------|
| # 0.0 μm          | 0.0 μm    | 0.0 μm        | ± 2.5 μm          | ±                           | 5.8E-02 μm |
| 24.1 μm           | 24.2 μm   | 0.1 μm        | ± 2.5 μm          | ±                           | 5.8E-02 μm |
| 53.5 μm           | 53.7 μm   | 0.2 μm        | ± 2.5 μm          | ±                           | 5.8E-02 μm |
| 126.0 μm          | 126 μm    | 0.0 μm        | ± 3.8 μm          | ±                           | 6.1E-02 μm |
| 255.0 μm          | 255 μm    | 0.0 μm        | ± 7.7 μm          | ±                           | 7.0E-02 μm |
| 493.0 μm          | 493 μm    | 0.0 μm        | ± 15 μm           | ±                           | 9.5E-02 μm |