

Patrón <i>(Standard)</i>	UBP <i>(UUT)</i>	Error <i>(Error)</i>	E.M.P <i>(Tolerance)</i>	Incertidumbre <i>(Uncertainty)</i>	
0.0000 mm	0.0000 mm	0.0000 µm	± 4 µm	±	5.8E+00 µm
5.0000 mm	5.0010 mm	1.0000 µm	± 4 µm	±	5.8E+00 µm
10.0000 mm	10.0014 mm	1.4000 µm	± 4 µm	±	5.8E+00 µm
15.0000 mm	15.0010 mm	1.0000 µm	± 4 µm	±	5.8E+00 µm
20.0000 mm	20.0012 mm	1.2000 µm	± 4 µm	±	5.8E+00 µm
25.0000 mm	25.0010 mm	1.0000 µm	± 4 µm	±	5.8E+00 µm