

Patrón <i>(Standard)</i>	UBP <i>(UUT)</i>	Error <i>(Error)</i>	Exactitud <i>(Accuracy)</i>	Incertidumbre <i>(Uncertainty)</i>	
0.0000 mm	0.0001 mm	0.1250 µm	± 30 µm	±	5.8E+00 µm
0.1000 mm	0.0973 mm	-2.7500 µm	± 30 µm	±	5.8E+00 µm
0.5000 mm	0.5008 mm	0.7500 µm	± 30 µm	±	5.8E+00 µm
1.0000 mm	1.0001 mm	0.1250 µm	± 30 µm	±	5.8E+00 µm
3.0000 mm	3.0020 mm	2.0000 µm	± 30 µm	±	5.8E+00 µm
5.0000 mm	5.0036 mm	3.6250 µm	± 30 µm	±	5.8E+00 µm
6.0000 mm	6.0009 mm	0.8750 µm	± 30 µm	±	5.8E+00 µm
9.0000 mm	8.9953 mm	-4.7500 µm	± 30 µm	±	5.8E+00 µm
10.0000 mm	9.9951 mm	-4.8750 µm	± 30 µm	±	5.8E+00 µm