

Patrón <i>(Standard)</i>	UBP <i>(UUT)</i>	Error <i>(Error)</i>	Exactitud <i>(Accuracy)</i>	Incertidumbre <i>(Uncertainty)</i>	
0.0000 mm	-0.0004 mm	-0.3750 μm	± 30 μm	±	5.8E+00 μm
0.1000 mm	0.0991 mm	-0.8750 μm	± 30 μm	±	5.8E+00 μm
0.5000 mm	0.5039 mm	3.8750 μm	± 30 μm	±	5.8E+00 μm
1.0000 mm	1.0054 mm	5.3750 μm	± 30 μm	±	5.8E+00 μm
3.0000 mm	3.0090 mm	9.0000 μm	± 30 μm	±	5.8E+00 μm
5.0000 mm	5.0106 mm	10.6250 μm	± 30 μm	±	5.8E+00 μm
6.0000 mm	6.0098 mm	9.7500 μm	± 30 μm	±	5.8E+00 μm
9.0000 mm	9.0044 mm	4.3750 μm	± 30 μm	±	5.8E+00 μm
10.0000 mm	10.0030 mm	3.0000 μm	± 30 μm	±	5.8E+00 μm