

Patrón <i>(Standard)</i>	UBP <i>(UUT)</i>	Error <i>(Error)</i>	Exactitud <i>(Accuracy)</i>	Incertidumbre <i>(Uncertainty)</i>	
0.0000 in	0.0000 in	0.0000 in	± 0.001 in	±	5.8E-04 in
0.0100 in	0.0099 in	-0.0001 in	± 0.001 in	±	5.8E-04 in
0.0500 in	0.0498 in	-0.0002 in	± 0.001 in	±	5.8E-04 in
0.1000 in	0.0997 in	-0.0003 in	± 0.001 in	±	5.8E-04 in
0.3000 in	0.2994 in	-0.0006 in	± 0.001 in	±	6.2E-04 in
0.5000 in	0.4995 in	-0.0005 in	± 0.001 in	±	6.8E-04 in
0.6000 in	0.5994 in	-0.0006 in	± 0.001 in	±	7.1E-04 in
0.9000 in	0.8990 in	-0.0010 in	± 0.001 in	±	8.5E-04 in
0.9800 in	0.9791 in	-0.0009 in	± 0.001 in	±	9.0E-04 in