

<b>Patrón (Standard)</b>	<b>UBP (UUT)</b>	<b>Error (Error)</b>	<b>E.M.P (Tolerance)</b>	<b>Incertidumbre (Uncertainty)</b>	
12.0000 in	12.0001 in	0.0001 in	± 0.0003 in	±	5.8E-04 in
12.2000 in	12.2003 in	0.0003 in	± 0.0004 in	±	5.8E-04 in
12.4000 in	12.4003 in	0.0003 in	± 0.0004 in	±	5.8E-04 in
12.6000 in	12.6003 in	0.0003 in	± 0.0004 in	±	5.8E-04 in
12.8000 in	12.8003 in	0.0003 in	± 0.0004 in	±	5.8E-04 in
13.0000 in	13.0003 in	0.0003 in	± 0.0004 in	±	5.8E-04 in