

Patrón (Standard)	UBP (UUT)	Error (Error)	E.M.P (Tolerance)	Incertidumbre (Uncertainty)	
13.0000 in	13.0001 in	0.0001 in	± 0.0004 in	±	5.8E-04 in
13.2000 in	13.2001 in	0.0001 in	± 0.0004 in	±	5.8E-04 in
13.4000 in	13.4001 in	0.0001 in	± 0.0004 in	±	5.8E-04 in
13.6000 in	13.6003 in	0.0003 in	± 0.0004 in	±	5.8E-04 in
13.8000 in	13.8001 in	0.0001 in	± 0.0004 in	±	5.8E-04 in
14.0000 in	14.0003 in	0.0003 in	± 0.0004 in	±	5.8E-04 in