

Patrón <i>(Standard)</i>	UBP <i>(UUT)</i>	Error <i>(Error)</i>	E.M.P <i>(Tolerance)</i>	Incertidumbre <i>(Uncertainty)</i>	
14.0000 in	14.0001 in	0.0001 in	± 0.0004 in	±	5.8E-04 in
14.2000 in	14.2001 in	0.0001 in	± 0.0004 in	±	5.8E-04 in
14.4000 in	14.4001 in	0.0001 in	± 0.0004 in	±	5.8E-04 in
14.6000 in	14.6001 in	0.0001 in	± 0.0004 in	±	5.8E-04 in
14.8000 in	14.8002 in	0.0002 in	± 0.0004 in	±	5.8E-04 in
15.0000 in	15.0003 in	0.0003 in	± 0.0004 in	±	5.8E-04 in