Patrón (Standard)	UBP (UUT)	Error (Error)	E.M.P (Tolerance)	Incertidumbre (Uncertainty)	
15.0000 in	15.0001 in	0.0001 in	$\pm 0.0004 in$	± 5.8E-04 in	
15.2000 in	15.2001 in	0.0001 in	± 0.0004 in	± 5.8E-04 in	
15.4000 in	15.4001 in	0.0001 in	$\pm 0.0004 in$	± 5.8E-04 in	
15.6000 in	15.6001 in	0.0001 in	$\pm 0.0004 in$	± 5.8E-04 in	
15.8000 in	15.8002 in	0.0002 in	$\pm 0.0004 in$	± 5.8E-04 in	
16.0000 in	16.0002 in	0.0002 in	± 0.0004 in	± 5.8E-04 in	