

Patrón <i>(Standard)</i>	UBP <i>(UUT)</i>	Error <i>(Error)</i>	E.M.P <i>(Tolerance)</i>	Incertidumbre <i>(Uncertainty)</i>	
16.0000 in	16.0001 in	0.0001 in	± 0.0004 in	±	5.8E-04 in
16.2000 in	16.2002 in	0.0002 in	± 0.0004 in	±	5.8E-04 in
16.4000 in	16.4002 in	0.0002 in	± 0.0004 in	±	5.8E-04 in
16.6000 in	16.6003 in	0.0003 in	± 0.0004 in	±	5.8E-04 in
16.8000 in	16.8003 in	0.0003 in	± 0.0004 in	±	5.8E-04 in
17.0000 in	17.0003 in	0.0003 in	± 0.0004 in	±	5.8E-04 in