

Patrón (Standard)	UBP (UUT)	Error (Error)	E.M.P (Tolerance)	Incertidumbre (Uncertainty)	
17.0000 in	17.0001 in	0.0001 in	± 0.0004 in	±	5.8E-04 in
17.2000 in	17.2001 in	0.0001 in	± 0.0004 in	±	5.8E-04 in
17.4000 in	17.4002 in	0.0002 in	± 0.0004 in	±	5.8E-04 in
17.6000 in	17.6003 in	0.0003 in	± 0.0004 in	±	5.8E-04 in
17.8000 in	17.8003 in	0.0003 in	± 0.0004 in	±	5.8E-04 in
18.0000 in	18.0003 in	0.0003 in	± 0.0004 in	±	5.8E-04 in