Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)	
0.0000 in	0.0002 in	0.0002 in	$\pm 0.001$ in	±	8.2E-04 in
0.0100 in	0.0100 in	0.0000 in	$\pm 0.001$ in	±	5.8E-04 in
0.0500 in	0.0500 in	0.0000 in	$\pm 0.001$ in	±	5.8E-04 in
0.1000 in	0.0997 in	-0.0003 in	$\pm 0.001$ in	±	5.9E-04 in
0.3000 in	0.2994 in	-0.0006 in	$\pm 0.001$ in	±	6.2E-04 in
0.5000 in	0.4997 in	-0.0003 in	$\pm 0.001$ in	±	6.8E-04 in
0.6000 in	0.5996 in	-0.0004 in	$\pm 0.001$ in	±	7.1E-04 in
0.9000 in	0.8991 in	-0.0009 in	$\pm 0.001$ in	±	8.5E-04 in
0.9800 in	0.9789 in	-0.0011 in	$\pm 0.001$ in	±	9.0E-04 in