

Clase de Exactitud (Accuracy Class): Clase II (Class II)

Valor UBP	Valor Patrón	Error	EMP	Incertidumbre (Uncertainty)		Valor UBP	Valor Patrón	Error	EMP	Incertidumbre (Uncertainty)	
m	m	mm	mm			m	m	mm	mm		
0 a 0.1	0.09975	0.250	± 0.6	±	4.3E-01 mm						
0 a 0.5	0.49961	0.390	± 0.8	±	4.3E-01 mm						
0 a 1	0.99951	0.490	± 1	±	4.3E-01 mm						
0 a 2	1.99978	0.220	± 1.4	±	4.8E-01 mm						
0 a 3	2.99940	0.600	± 1.8	±	4.8E-01 mm						
0 a 4	3.99923	0.770	± 2.2	±	4.8E-01 mm						
0 a 5	4.99957	0.430	± 2.6	±	4.8E-01 mm						
0 a 6	5.99958	0.420	± 3	±	3.6E-01 mm						
0 a 7	6.99894	1.060	± 3.4	±	3.6E-01 mm						
0 a 8	7.99904	0.960	± 3.8	±	3.6E-01 mm						