

Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)	
1 mm	1.00 mm	0 µm	± 30 µm	±	5.8E+00 µm
2 mm	2.00 mm	0 µm	± 30 µm	±	5.8E+00 µm
3 mm	3.00 mm	0 µm	± 30 µm	±	5.8E+00 µm
4 mm	4.00 mm	0 µm	± 30 µm	±	5.8E+00 µm
5 mm	5.00 mm	0 µm	± 30 µm	±	5.8E+00 µm
6 mm	6.00 mm	0 µm	± 30 µm	±	5.8E+00 µm
7 mm	7.00 mm	0 µm	± 30 µm	±	5.8E+00 µm
8 mm	8.00 mm	0 µm	± 30 µm	±	5.8E+00 µm
9 mm	9.00 mm	0 µm	± 30 µm	±	5.8E+00 µm
10 mm	10.00 mm	0 µm	± 30 µm	±	5.8E+00 µm