

<b>Patrón (Standard)</b>	<b>UBP (UUT)</b>	<b>Error (Error)</b>	<b>E.M.P (Tolerance)</b>	<b>Incertidumbre (Uncertainty)</b>	
20.0000 in	20.0001 in	0.0001 in	± 0.0005 in	±	5.8E-04 in
20.2000 in	20.2001 in	0.0001 in	± 0.0005 in	±	5.8E-04 in
20.4000 in	20.4003 in	0.0003 in	± 0.0005 in	±	5.8E-04 in
20.6000 in	20.6003 in	0.0003 in	± 0.0005 in	±	5.8E-04 in
20.8000 in	20.8001 in	0.0001 in	± 0.0005 in	±	5.8E-04 in
21.0000 in	21.0001 in	0.0001 in	± 0.0005 in	±	5.8E-04 in