

Patrón <i>(Standard)</i>	UBP <i>(UUT)</i>	Error <i>(Error)</i>	E.M.P <i>(Tolerance)</i>	Incertidumbre <i>(Uncertainty)</i>	
21.0000 in	21.0001 in	0.0001 in	± 0.0005 in	±	5.8E-04 in
21.2000 in	21.2003 in	0.0003 in	± 0.0005 in	±	5.8E-04 in
21.4000 in	21.4003 in	0.0003 in	± 0.0005 in	±	5.8E-04 in
21.6000 in	21.6003 in	0.0003 in	± 0.0005 in	±	5.8E-04 in
21.8000 in	21.8003 in	0.0003 in	± 0.0005 in	±	5.8E-04 in
22.0000 in	22.0003 in	0.0003 in	± 0.0005 in	±	5.8E-04 in