Patrón (Standard)	UBP (UUT)	Error (Error)	E.M.P (Tolerance)	Incertidumbre (Uncertainty)
22.0000 in	22.0003 in	0.0003 in	$\pm 0.0005 \text{ in}$	± 5.8E-04 in
22.2000 in	22.2003 in	0.0003 in	± 0.0005 in	± 5.8E-04 in
22.4000 in	22.4003 in	0.0003 in	± 0.0005 in	± 5.8E-04 in
22.6000 in	22.6003 in	0.0003 in	± 0.0005 in	± 5.8E-04 in
22.8000 in	22.8003 in	0.0003 in	± 0.0005 in	± 5.8E-04 in
23.0000 in	23.0003 in	0.0003 in	$\pm 0.0005 \text{ in}$	± 5.8E-04 in