

<b>Patrón (Standard)</b>	<b>UBP (UUT)</b>	<b>Error (Error)</b>	<b>E.M.P (Tolerance)</b>	<b>Incertidumbre (Uncertainty)</b>	
23.0000 in	23.0001 in	0.0001 in	± 0.0005 in	±	5.8E-04 in
23.2000 in	23.2001 in	0.0001 in	± 0.0005 in	±	5.8E-04 in
23.4000 in	23.4003 in	0.0003 in	± 0.0005 in	±	5.8E-04 in
23.6000 in	23.6002 in	0.0002 in	± 0.0005 in	±	5.8E-04 in
23.8000 in	23.8001 in	0.0001 in	± 0.0005 in	±	5.8E-04 in
24.0000 in	24.0001 in	0.0001 in	± 0.0005 in	±	5.8E-04 in