

Clase de Exactitud (Accuracy Class): Clase III (Class III)

Valor UBP	Valor Patrón	Error	EMP	Incertidumbre	Valor UBP	Valor Patrón	Error	EMP	Incertidumbre
m	m	mm	mm	(Uncertainty)	m	m	mm	mm	(Uncertainty)
0 a 1	0.99859	1.410	± 2	± 4.3E-01 mm					
0 a 2	2.00068	-0.680	± 2.8	± 4.8E-01 mm					
0 a 3	2.99981	0.190	± 3.6	± 4.8E-01 mm					
0 a 4	4.00019	-0.190	± 4.4	± 4.8E-01 mm					
0 a 5	4.99991	0.090	± 5.2	± 4.8E-01 mm					
0 a 6	6.00044	-0.440	± 6	± 3.6E-01 mm					
0 a 7	7.00015	-0.150	± 6.8	± 3.6E-01 mm					
0 a 8	8.00071	-0.710	± 7.6	± 3.6E-01 mm					
0 a 9	9.00031	-0.310	± 8.4	± 3.6E-01 mm					
0 a 10	10.00107	-1.070	± 9.2	± 3.6E-01 mm					
0 a 11	11.00044	-0.440	± 10	± 3.6E-01 mm					
0 a 12	12.00115	-1.150	± 10.8	± 3.6E-01 mm					
0 a 13	13.00059	-0.590	± 11.6	± 3.6E-01 mm					
0 a 14	14.00113	-1.130	± 12.4	± 3.6E-01 mm					
0 a 15	15.00103	-1.030	± 13.2	± 3.6E-01 mm					
