

Patrón <i>(Standard)</i>	UBP <i>(UUT)</i>	Error <i>(Error)</i>	Exactitud <i>(Accuracy)</i>	Incertidumbre <i>(Uncertainty)</i>	
0.0000 mm	0.0003 mm	0.3333 μm	± 0.01 μm	±	5.8E+00 μm
0.1000 mm	0.0985 mm	-1.5000 μm	± 0.01 μm	±	5.8E+00 μm
0.5000 mm	0.4983 mm	-1.6667 μm	± 0.01 μm	±	5.8E+00 μm
1.0000 mm	0.9980 mm	-2.0000 μm	± 0.01 μm	±	5.8E+00 μm
2.0000 mm	1.9960 mm	-4.0000 μm	± 0.01 μm	±	5.8E+00 μm
2.5000 mm	2.4930 mm	-7.0000 μm	± 0.01 μm	±	5.8E+00 μm
3.0000 mm	2.9940 mm	-6.0000 μm	± 0.01 μm	±	5.8E+00 μm
4.0000 mm	3.9955 mm	-4.5000 μm	± 0.01 μm	±	5.8E+00 μm
5.0000 mm	4.9973 mm	-2.6667 μm	± 0.01 μm	±	5.8E+00 μm