

Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
0.0000 mm	0.0003 mm	0.3333 μm	$\pm 30 \mu\text{m}$	$\pm 5.8\text{E}+00 \mu\text{m}$
0.1000 mm	0.0985 mm	-1.5000 μm	$\pm 30 \mu\text{m}$	$\pm 5.8\text{E}+00 \mu\text{m}$
0.5000 mm	0.4983 mm	-1.6667 μm	$\pm 30 \mu\text{m}$	$\pm 5.8\text{E}+00 \mu\text{m}$
1.0000 mm	0.9980 mm	-2.0000 μm	$\pm 30 \mu\text{m}$	$\pm 5.8\text{E}+00 \mu\text{m}$
2.0000 mm	1.9960 mm	-4.0000 μm	$\pm 30 \mu\text{m}$	$\pm 5.8\text{E}+00 \mu\text{m}$
2.5000 mm	2.4930 mm	-7.0000 μm	$\pm 30 \mu\text{m}$	$\pm 5.8\text{E}+00 \mu\text{m}$
3.0000 mm	2.9940 mm	-6.0000 μm	$\pm 30 \mu\text{m}$	$\pm 5.8\text{E}+00 \mu\text{m}$
4.0000 mm	3.9955 mm	-4.5000 μm	$\pm 30 \mu\text{m}$	$\pm 5.8\text{E}+00 \mu\text{m}$
5.0000 mm	4.9973 mm	-2.6667 μm	$\pm 30 \mu\text{m}$	$\pm 5.8\text{E}+00 \mu\text{m}$