

Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
0.0000 mm	0.0005 mm	0.5000 μm	$\pm 30 \mu\text{m}$	$\pm 5.8\text{E}+00 \mu\text{m}$
0.1000 mm	0.0992 mm	-0.8333 μm	$\pm 30 \mu\text{m}$	$\pm 5.8\text{E}+00 \mu\text{m}$
0.5000 mm	0.4993 mm	-0.6667 μm	$\pm 30 \mu\text{m}$	$\pm 5.8\text{E}+00 \mu\text{m}$
1.0000 mm	1.0020 mm	2.0000 μm	$\pm 30 \mu\text{m}$	$\pm 5.8\text{E}+00 \mu\text{m}$
2.0000 mm	2.0030 mm	3.0000 μm	$\pm 30 \mu\text{m}$	$\pm 5.8\text{E}+00 \mu\text{m}$
2.5000 mm	2.5005 mm	0.5000 μm	$\pm 30 \mu\text{m}$	$\pm 5.8\text{E}+00 \mu\text{m}$
3.0000 mm	3.0023 mm	2.3333 μm	$\pm 30 \mu\text{m}$	$\pm 5.8\text{E}+00 \mu\text{m}$
4.0000 mm	4.0033 mm	3.3333 μm	$\pm 30 \mu\text{m}$	$\pm 5.8\text{E}+00 \mu\text{m}$
5.0000 mm	5.0008 mm	0.8333 μm	$\pm 30 \mu\text{m}$	$\pm 5.8\text{E}+00 \mu\text{m}$