

Patrón <i>(Standard)</i>	UBP <i>(UUT)</i>	Error <i>(Error)</i>	Exactitud <i>(Accuracy)</i>	Incertidumbre <i>(Uncertainty)</i>	
0.0000 mm	0.0005 mm	0.5000 µm	± 0.001 µm	±	5.8E+00 µm
0.1000 mm	0.0992 mm	-0.8333 µm	± 0.001 µm	±	5.8E+00 µm
0.5000 mm	0.4993 mm	-0.6667 µm	± 0.001 µm	±	5.8E+00 µm
1.0000 mm	1.0020 mm	2.0000 µm	± 0.001 µm	±	5.8E+00 µm
2.0000 mm	2.0030 mm	3.0000 µm	± 0.001 µm	±	5.8E+00 µm
2.5000 mm	2.5005 mm	0.5000 µm	± 0.001 µm	±	5.8E+00 µm
3.0000 mm	3.0023 mm	2.3333 µm	± 0.001 µm	±	5.8E+00 µm
4.0000 mm	4.0033 mm	3.3333 µm	± 0.001 µm	±	5.8E+00 µm
5.0000 mm	5.0008 mm	0.8333 µm	± 0.001 µm	±	5.8E+00 µm