

Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)	
0.0000 mm	0.0005 mm	0.5000 µm	± 0.01 µm	±	5.8E+00 µm
0.1000 mm	0.0992 mm	-0.8333 µm	± 0.01 µm	±	5.8E+00 µm
0.5000 mm	0.4993 mm	-0.6667 µm	± 0.01 µm	±	5.8E+00 µm
1.0000 mm	1.0020 mm	2.0000 µm	± 0.01 µm	±	5.8E+00 µm
2.0000 mm	2.0030 mm	3.0000 µm	± 0.01 µm	±	5.8E+00 µm
2.5000 mm	2.5005 mm	0.5000 µm	± 0.01 µm	±	5.8E+00 µm
3.0000 mm	3.0023 mm	2.3333 µm	± 0.01 µm	±	5.8E+00 µm
4.0000 mm	4.0033 mm	3.3333 µm	± 0.01 µm	±	5.8E+00 µm
5.0000 mm	5.0008 mm	0.8333 µm	± 0.01 µm	±	5.8E+00 µm