

Patrón <i>(Standard)</i>	UBP <i>(UUT)</i>	Error <i>(Error)</i>	Exactitud <i>(Accuracy)</i>	Incertidumbre <i>(Uncertainty)</i>	
0.0000 in	0.0000 in	0.0000 in	± 0.001 in	±	5.8E-04 in
0.0100 in	0.0098 in	-0.0002 in	± 0.001 in	±	5.9E-04 in
0.0500 in	0.5000 in	0.4500 in	± 0.001 in	±	6.8E-04 in
0.0900 in	0.0901 in	0.0001 in	± 0.001 in	±	5.9E-04 in
0.1000 in	0.1003 in	0.0003 in	± 0.001 in	±	5.9E-04 in
0.1100 in	0.1101 in	0.0001 in	± 0.001 in	±	5.8E-04 in
0.1500 in	0.1501 in	0.0001 in	± 0.001 in	±	6.0E-04 in
0.1900 in	0.1898 in	-0.0002 in	± 0.001 in	±	6.0E-04 in
0.2000 in	0.2002 in	0.0002 in	± 0.001 in	±	6.1E-04 in