

| Patrón <i>(Standard)</i> | UBP <i>(UUT)</i> | Error <i>(Error)</i> | Exactitud <i>(Accuracy)</i> | Incertidumbre <i>(Uncertainty)</i> | |
|--------------------------|------------------|----------------------|-----------------------------|------------------------------------|------------|
| 0.0000 in | 0.0000 in | 0.0000 in | ± 0.001 in | ± | 6.0E-04 in |
| 0.0100 in | 0.0102 in | 0.0002 in | ± 0.001 in | ± | 6.0E-04 in |
| 0.0500 in | 0.0497 in | -0.0003 in | ± 0.001 in | ± | 6.1E-04 in |
| 0.0900 in | 0.0900 in | 0.0000 in | ± 0.001 in | ± | 7.0E-04 in |
| 0.1000 in | 0.1002 in | 0.0002 in | ± 0.001 in | ± | 7.0E-04 in |
| 0.1100 in | 0.1098 in | -0.0002 in | ± 0.001 in | ± | 6.2E-04 in |
| 0.1500 in | 0.1499 in | -0.0001 in | ± 0.001 in | ± | 1.0E-03 in |
| 0.1900 in | 0.1901 in | 0.0001 in | ± 0.001 in | ± | 8.4E-04 in |
| 0.2000 in | 0.2003 in | 0.0003 in | ± 0.001 in | ± | 6.2E-04 in |